



<b>Organisation Chain :</b>	Department of Pharmaceuticals  National Institute of Pharmaceutical Education and Research Hajipur
<b>Tender ID :</b>	2022_MCF_669015_1
<b>Tender Ref No :</b>	Tender No NIPER HJP/S and P/11/GC-MS QQQ/2022-23
<b>Tender Title :</b>	Tender No NIPER HJP/S and P/11/GC-MS QQQ/2022-23
<b>Corrigendum Type :</b>	Other

**Corrigendum Document Details**

Corr.No.	Corrigendum Title	Corrigendum Description	Published Date	Document Name	Doc Size(in KB)
1	Amendment of NIPER GC-MS QQQ Tender	As per tender document and Amended specification	26-Jul-2022 04:52 PM	<a href="#">GCMS Amendment.pdf</a>	253.48



# राष्ट्रीय औषधीय शिक्षा एवं अनुसंधान संस्थान—हाजीपुर

NATIONAL INSTITUTE OF PHARMACEUTICAL EDUCATION AND RESEARCH (NIPER) - HAJIPUR

(Department of Pharmaceuticals, Ministry of Chemicals & Fertilizers, Govt. of India)

Export Promotions Industrial Park (EPIP), Industrial Area, Hajipur

Dist: Vaishali, State: Bihar, India, PIN: 844102 Website: [www.niperhajipur.ac.in](http://www.niperhajipur.ac.in)

Amendment of NIPER GC-MS QQQ Tender

(Tender reference No - NIPER HJP/S and P/11/GC-MS QQQ/2022-23

Tender Id -2022\_MCF\_669015\_1 dated 12.07.2022

Sl. No.	Previous Specification	Amended specification
1	(Section E, Point no 19 and 20) The sensitivity of the MS/MS system should be as follows: EI MRM , 1 µl of 100 fg/µl OFN (m/z 272 □ 222), S/N 20000:1 or better Detection limit should be 0.5 fg or less octafluoronaphthalene (OFN)with 99% confidence level.	The sensitivity of the MS/MS system should be as follows: EI MRM , 1 µl of 100 fg/µl OFN (m/z 272 □ 222), S/N <b>30000:1</b> or better Detection limit should be 0.5 fg or less octafluoronaphthalene (OFN)with 99% confidence level.
2	(section E, point no 19) No additional Points earlier	Additional Points a. For CI MRM, S/N for 1pg - 5pg Benzophenone S/N 2000:1 or better b. NCI SIM/full scan, S/N 100fg -1pg OFN (m/z 272), S/N 2000:1 or better
3	(Section A, Point no-5) Offered system should be latest model	Most upgraded model to be quoted for both GC and MS units.
4	(Section J, Other requirements)	Additional point Preventive and maintenance kits and accessories required for maintenance will be provided by the vendor

Other terms and conditions remains unchanged.

S/d  
Registrar I/C  
(NIPER-Hajipur)